



BELZ CALIBRATION LAB

(A DIVISION OF BELZ INSTRUMENTS PVT. LTD.)

5L-123, N.I.T. Faridabad-121001. (Haryana) # 2415646, 2414855 Fax : 4055855

E-mail : info@belzcallab.com www.belzcallab.com



Calibration Certificate

FORMAT NO. QS-22F

Certificate no.: 6105766	Service Request No. : 2013/2670	Page : 01
Issue Date : 14/11/2013	Service Request Date: 13/11/2013	Total Pages : 02
	Calibration Date : 14/11/2013	

Name and Address of Client	Instrument Details
MRM PROCOM PRIVATE LIMITED PLOT NO.20-21, INDUSTRIAL ESTATE, SECTOR-59, PHASE-II, FARIDABAD-121004, HARYANA.	Name : VIF METER Make : PROCOM Sr.No. : D2013.10813 ID.No. : ---- Range/Size : AS PER INSTRUMENT Least Count : AS PER RANGE Accuracy : --- Location : ---

Ref. Std./Cal.Procedure : CI(ES)-13 Results Reported at Temp(°C): 25±3 Relative Humidity(%) : 35 TO 70	Condition of Instrument Visual Inspection: OK Zero Error : NIL Repeatability : ---
--	--

Standard Equipment used for calibration

Name	9100 FLUKE CALIBRATION SYSTEM	ACCUCHEK LT+
Sr.No.	941254428	ALP00258
Model	9100	LT+
Make	FLUKE	SECURE METERS LIMITED
Calibrated By	ERTL(N)	YMPL
Certificate No.	ERTL(N)/90(4)2013-14/A0199	YMPL/218856/47486
Validity of Calibration	27/06/2014	10/10/2014

Notes:

- This Certificate refers only to the particular item submitted for calibration.
- This Certificate shall not be reproduced, except in full, without the written permission of Chief Executive Belz Instruments Pvt. Ltd., FARIDABAD.
- Results Reported are valid at the time of and under the stated conditions of measurement.
- Laboratory Standards are traceable to National Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

Calibrated by: VARUN DUBEY
Calibration Engg.

Approved by: N.K. SHARMA
Quality Manager



BELZ CALIBRATION LAB

(A DIVISION OF BELZ INSTRUMENTS PVT. LTD.)

5L-123, N.I.T. Faridabad-121001. (Haryana) # 2415646, 2414855 Fax : 4055855
E-mail : info@belzcallab.com www.belzcallab.com



Calibration Certificate

Certificate no. : 6105766	Service Request No. : 2013/2670	Page : 02
Date of issue : 14/11/2013	Service Request Date : 13/11/2013	Total Pages : 02
	Calibration Date : 14/11/2013	

Calibration Results

RANGE	Std.Readings in (Volt)	Unit Under Calib.in (Volt)			UNCERTAINTY AT 95% CONFIDENCE LEVEL(k=2)
		R	Y	B	
VOLTAGE (L-N)	50.00	50.12	50.13	50.12	+/-1%
	100.00	100.1	100.2	100.1	
	230.00	230.0	230.2	230.1	
VOLTAGE (L-L)	400.00	400.2	400.3	400.3	+/-1%
	440.00	440.3	440.4	440.2	
CURRENT	0.500	0.504	0.501	0.500	+/-1%
	1.000	1.002	1.003	1.001	
	3.000	3.004	3.004	3.002	
	5.000	4.996	4.998	4.994	
FREQUENCY	45.000	45.01			+/-0.2%
	50.000	50.02			
	55.000	55.01			

Remarks : ---

Calibration valid upto : 13/11/2014

- Notes:
- This Certificate refers only to the particular item submitted for calibration.
- This Certificate shall not be reproduced, except in full, without the permission of Chief Executive BELZ INSTRUMENTS PVT. LTD. FARIDABAD
- Results Reported are valid at the time of and under the stated conditions of measurement.
- Laboratory Standard are traceable to National Standards.
- Certificate issued for Mass, Weighing Balance, Volumetric equipments, Steel scale & Measuring Tapes are for scientific use only & not for Trade/Commercial use.

Calibrated by: VARUN DUBEY
Calibration Engg.

Approved By: R.K. KHERRA
Quality Manager

